

Tyrosine

CHROMATOGRAPHIC CONDITIONS

Column: CROWNPAK® CR-I (+) [3.0 x 150 mm, 5 μ m]

Part No: 53784

Mobile phase: aq. Perchloric acid (pH=1.0) : ACN (80/20, v/v)


Flowrate: 0.3 mL/min

Column Oven Temperature: 25 °C

UV Detection: 276 nm

Diluent: Mobile phase

Sample Concentration: 0.3 mg/mL



SEPARATION PERFORMANCE

t_{R1} : 4.0

N_1 : 6167

T_1 : 1.4

t_{R2} : 7.5

N_2 : 5564

T_2 : 1.7

R_s : 11.4

